

**Pogorielov M. In-vitro evaluation of new chitosan-based materials / M.Pogorielov, O. Kalinkevich, O. Pogorielova, G. Tkach / J of Biomedical and Clinical Research. - 2014. - V. 7(1). - P.17**

### **In-vitro evaluation of new chitosan-based materials**

\*Pogorielov M.V., <sup>1</sup>Kalinkevich O.V., Pogorielova O.S., Tkach G.F.

Sumy State University, Medical Institute, 31 Sanatornaya Street, Sumy 40007, Ukraine

<sup>1</sup>Institute for Applied Physics, National Academy of Sciences of Ukraine, 58 Petropavlovskaya Street, 40000, Sumy, Ukraine

\*Corresponding author:

[pogorelov\\_max@mail.ru](mailto:pogorelov_max@mail.ru)

+38 (0542) 9005448

The chitosan is a promising material for biomedical application due to nontoxicity, biodegradable, biofunctional, biocompatible properties in addition to having antimicrobial characteristics. We synthesized a series of new one step co-preparation materials based on chitosan for bone, skin and dura mater replacement. The best way to selection materials with satisfactory properties is in-vitro evaluation on cell cultures.

We used 3 different types of materials – chitosan films (200, 500 and 700 kDa), chitosan sponge and composite (chitosan-chitin and chitosan-apatite). All experiment was conducted in Department of Materials Science and Engineering, University of Sheffield, UK. For chitosan films and sponge and chitosan-chitin composite we used Murine long bone osteoblast (MLO) and for chitosan film and chitosan-apatite composite - Human oral mucosal fibroblasts (OF). We seeded 20 000 cells on each scaffold and evaluated cell viability

(material toxicity) on 1, 4, 7 and 14 days for MLO and on 1, 4 and 7 days for OF. Also we evaluated collagen production on the end time-point of experiment.

Results show that all materials are non-toxic and allow for cells growth and stimulate collagen secretion by the osteoblasts and fibroblasts. But the chitosan with high molecular weight (700 kDa) and chitosan-chitin composite have better biological response.